


EKG cases


DCS' Arytmikursus

Sandbjerg, 1.-2. september 2022

Case 1




Case 2



Case 3

10 mm/mV

5 mm/mV



Case 4


10 mm/mV 25 mm/s Filter: H50 d 100 Hz

10 mm/mV




Case 5

5.




Patient:




Case 6

6a.






6b.



Case 7


CASE 7



Case 8

10 mm/mV

10 mm/mV



25 mm/s

F50 SSF SBS 5a 01-AUG-09 11:36:08 Skejby Sygehus Afd.B1


AT-2plus 3.01 C

QT 362 ms R (V5) 1.28 mV
QTc 416 ms Sokol. 4.37 mV

After 7.5 mg Verapoc

10 mm/mV

10 mm/mV



25 mm/s

F50

Sa 01 AUG 09 18:24:23

SKR by Sygarius Rfd. B

AT-Vplus C 1.23

Case 9



Initial Interrogation Report

Pacemaker Model: Medtronic Sensia SESR01 Serial Number: NWR695974

Date of Interrogation: 08/05/15 4:46 PM

Pige, 11 år

Permanent Parameters

Modes

Mode VVI

Rates

Lower Rate 60 ppm

ADL Rate 95 ppm

Refractory/Blanking

Ventricular Refractory 330 ms

Rate Response

Optimization On
 ADL Response 3
 Exertion Response 3
 ADLR Percent 2.0%
 Activity Threshold Medium/Low
 Activity Acceleration 30 sec
 Activity Deceleration Exercise
 High Rate Percent 0.2%
 ADL Rate Setpoint 54
 Upper Sensor Rate Setpoint 154

Ventricular Lead

Amplitude 2.000 V
 Pulse Width 0.40 ms
 Sensitivity 4.00 mV
 Sensing Assurance On
 Pace Polarity Unipolar
 Sense Polarity Unipolar
 Lead Monitor Monitor Only
 Maximum Impedance 4,000 ohms
 Minimum Impedance 200 ohms
 Monitor Sensitivity 8
 Capture Management Adaptive
 Amplitude Margin 2x
 Min. Adapted Amplitude 2,000 V
 Capture Test Frequency Day at Rest
 Acute Phase Off
 Acute Phase Complete 09/04/11

Additional Features




Sleep Off
 Single Chamber Hysteresis Off
 Transtelephonic Monitor Off
 Extended Telemetry Off
 Extended Marker Standard
 Implant Detection Off/Complete

Ventricular High Rate Episodes


Detection Rate 180 ppm
 Detection Beats 5 beats
 Termination Beats 5 beats
 Episode Collection Method Rolling

Selectable Diagnostic

Chronic Lead Trend On
 High Rate Detail
 Include Refractory Senses? Include
 EGM Type EGM
 EGM Allocation 4 for 2/2 secs
 EGM Timeout 8 weeks




Pige, 11 år



Case 10

4.




Case 11


11.



11.




Case 12



Case 13

CASE 13




Case 14

CASE 14

QRS 65
T 220

- ABNORMAL ECG -

Ikke bekreftet diagnose.




25 mm/s 1.0 mm/mV 0.15 Hz - 40 Hz HP709 13337

Case 15

CASE 15

200 bpm




0013
0017503
2013
1051

Case 16

CASE 16a




CASE 16b




CASE 16c


Adenosin




Case 17



17.




His




Case 18

CASE 18a



CASE 18b




Side 1

B2

Slagelse Sygehus

CASE 18c




Case 19

CASE 19a



CASE 19b



CASE 19c




Case 20


CASE 20a



CASE 20b




CASE 20c



CASE 20d



Case 21



Episode: Non-sustained (468 bpm / 128 ms)

VT/VF Episode 2 of 2




Page 2 of 3

17 Oct 2015 21:25

1: V Sense Amp AutoGain (0,4 mm/mV)
2: Discrimination AutoGain (1,8 mm/mV)

3: Markers

Sweep Speed: 25 mm/s



Episode: VF (363 min⁻¹ / 165 ms) (Continued)


VT/VF Episode 22 of 35

Page 3 of 4

18 Oct 2015 17:21

- 1: V Sense Amp AutoGain (0.5 mm/mV)
- 2: Discrimination AutoGain (3.6 mm/mV)
- 3: Markers

Sweep Speed: 25 mm/s



VT/VF Episode 22 of 35





Page 4 of 4

Episode: VF (363 min⁻¹ / 165 ms) (Continued)


18 Oct 2015 17:21

- 1: V Sense Amp AutoGain (0.5 mm/mV)
- 2: Discrimination AutoGain (3.6 mm/mV)
- 3: Markers

Sweep Speed: 25 mm/s



Case 22



12 årig dreng

ADHD

Planlægger Ritalin behandling


Der tages EKG

Case 23

A


V

Marker




Case 24

CASE 24




Case 25




Case 26

CASE 26




Case 27

CASE 27a




DR
10/10/10
200
1000

CASE 27b



CASE 27c

I forbindelse med trappegang til 2.sal bliver pt. udtalt dyspnøisk og svimmel




Case 28

CASE 28a


25 mm/s Filter 100 Hz H 50 d Filter 35 Hz



CASE 28b




CASE 28c




Case 29

CASE 29



CASE 29a




Case 30

CASE 30

1 sek


(a)



a) DDD-PM
Paced AV=280 ms.
Unipolar atrie-elektrode.
Unipolar ventrikel-elektrode

(b)

1 sek



b) DDD-PM
Paced AV=300 ms
Bipolar atrie-elektrode
Unipolar ventrikel-elektrode

(c)


1 sek



c) DDD-PM
Paced AV=300 ms
Bipolar atrie-elektrode
Unipolar ventrikel-elektrode


Case 31

CASE 31a




CASE 31b

Ver. 6.359281 / 2.159481




CASE 31c



Case 32

CASE 32




Case 33

CASE 33a




CASE 33b




Case 34

CASE 34



Case 35

CASE 35




Case 36


CASE 36



Case 37

37. Baseline-EKG






11/15/11

20:57:34(1)



Case 38


CASE 38



Case 39

CASE 39


ca. 280 bpm




25 mm/s 10 mm/mV 100Hz 50Hz MEGACART

Case 40


CASE 40



Case 41



25 mm/s 10 mm/mV



25 mm/s 10 mm/mV



Case 42

PRM02

I

II

III

AVR

AVL

AVF

V1

V2


V3

V4

V5

V6

10 mm/mV 10 mm/s



PRM02

I

II

III

AVR

AVL

AVF

V1

V2

V3

V4



V5

V6

10 mm/mV 25 mm/s



Case 43



Case 44

FJ02

— 1 sek

I

II

III

AVR

AVL

AVF

V1

V2


V3

V4


V5

V6

10 mm/mV 10 mm/s



1 sek



Case 45




Tekniker: siih
Test ind:

Henvist af:


Ubekræftet

Kommentar:




Case 46


CASE 46a



CASE 46b




CASE 46c



25 mm/s
10 mm/mV

Case 47

CASE 47a




Case 48

CASE 48

PVC PVC PVC PVC, V1ACH PVC

05:34 EXERCISE.



Case 49

CASE 49

10 mm/mV



25 mm/s


0.05-35 Hz F50

12:03:43

A. Schinopoulos
2016
TSJ



Case 50

CASE 50



Case 51


CASE 51




Case 52

CASE 52


Continuous strip 16:26 Avg 46bpm Ch2




Continuous strip




Continuous strip




Continuous strip




Continuous strip




Continuous strip



Continuous strip



Continuous strip



Case 53

CASE 53a

10 mm/mV


25 mm/s

Filter 100 Hz H 50 d

10 mm/mV



CASE 53b



Case 54


CASE 54a



CASE 54b





CASE 54c



Case 55

CASE 55




Case 56

Kommentar:

Henvist af:

Ubekræftet




25mm/s 10mm/mV 150Hz 8.0.1 12SL 241 HD Vogn-ID: 3

EID: EDT: BESTILLING:



25mm/s 10mm/mV 150Hz 8.0.1 I2SL 241 HD Vogn-ID: I


EID: EDT: BESTILLING:






Case 57

Type	ATP Seq	Shocks	Success	ID#	Date	Time hh:mm	Duration hh:mm:ss	Avg bpm V	Max bpm V	Activity at Onset
VF	1	35J	Yes	36	18-Dec-2013	14:31	:01:08	286	---	Active

• V-V VF = 300 ms FVT = 240 ms



60 årig mand, kendt med atrieflimren. Non-IHD kardiomyopati, primær profylaktisk ICD. Hård fysisk arbejde i forbindelse med bilreparation. Besvimer



Episode Summary

Initial Type	VF (spontaneous)
Duration	1.1 min
V. Max Rate	---
V. Median	261 bpm (230 ms)
Activity at onset	Active, Sensor = 83 bpm
Last Therapy	VF Rx1: Defib, Successful

Initial VT/VF Detection

Withheld By	None
--------------------	------

Therapies Delivered Charge Ohms Energy

VF Rx 1 Burst	During Charging			
VF Rx 1 Defib	34.6 J	8.06 sec	48 ohms	0.0 - 35 J
Termination				

Wavelet Measurements Prior to Initial VT/VF Detection

Wavelet Result: Wavelet not applied; interval too fast or too slow
 Template Status: OK

- 8. No Match 58 %
- 7. No Match --- Interval too fast
- 6. No Match --- Interval too fast
- 5. Match 70 %
- 4. No Match 43 %
- 3. No Match --- Interval too fast
- 2. No Match --- Interval too fast
- 1. No Match --- Interval too fast
- 0. Detection

Parameter Summary

Mode	VVI	Lower Rate	40 bpm
Detection		Rates	Therapies
VF	On	>200 bpm	ATP During Charging, 35J x 6
FVT	via VF	200-250 bpm	Burst(2), Ramp(2), 35J x 4
VT	OFF		All Rx Off
Enhancements On: Wavelet, TWave, Noise			

Parameter Settings Initial Redetect V. Interval (Rate)

VF	On	30/40	12/16	300 ms (200 bpm)
FVT	via VF			240 ms (250 bpm)
VT	Off	16	12	
Monitor	Off	20		

Wavelet

Wavelet	On, Match = 70 %
Template	05-Oct-2012, Auto = On
SVT V. Limit	260 ms

Other Enhancements

Stability	Off
Onset	Off
High Rate Timeout	
VF Zone Only	Off
TWave	On
RV Lead Noise	On


EGM Source Range Sensitivity

EGM1	RVtip to RVring	+/- 8 mV	RV	0.3 mV
EGM2 (Wavelet)	Can to RVcoil	+/- 8 mV		

Case 58

25 mm/s Filter 35 Hz H 50 d


10 mm/mV




Case 59

59a

Adenosin iv.



59b (cont.)




10.0 mm/mV

MAC35 010A.1

0.16-150 Hz


25.0 mm/s

59c (cont.)




Case 60

10 mm/mV 25 mm/s filter:150 Hz



Case 61




Puls er over 150 - ❤️ 177 SPM i gennemsnit

Dette ekg blev ikke undersøgt for atrieflimren, fordi din puls var over 150 SPM.


Hvis du gentagne gange får dette resultat, eller hvis du føler dig utilpas, bør du tale med din læge.

Rapporterede symptomer

- Hurtig puls, uregelmæssig hjerterytme eller hjertebanken
- Træthed
- Andet



25 mm/s, 10 mm/mV, Elektrode I, 512 Hz, iOS 15.3.1, watchOS 8.4.2, Watch6,1, Algoritmeversion 2 – Bølgeformen svarer til et ekg med 1 afledning (Lead I ECG). Se brugermanualen for at få flere oplysninger.



Case 62

Event detaljer: Tid: 29. juni 2022 17:34:08 Type: Pause Varighed: 5sek Hjertefrekvens i event af 7 slag: Lavest:17, Middel:56, Højest:205	Pause event grænseværdi: 3sek Kommentar: _____
---	---



Event detaljer: Tid: 29. juni 2022 17:40:11 Type: Pause Varighed: 3sek Hjertefrekvens i event af 9 slag: Lavest:17, Middel:74, Højest:225	Pause event grænseværdi: 3sek Kommentar: _____
---	---





Event detaljer: Tid: 29. juni 2022 18:36:07 Type: Pause Varighed: 4sek Hjertefrekvens i event af 8 slag: Lavest:19, Middel:65, Højest:132	Pause event grænseværdi: 3sek Kommentar: _____
---	---








Case 63













25 mm/s 10 mm/mV

Case 64




Case 65

65 a



65 b



SEMA-200 271.0 /B11.07407

(MS-2015 310)

Side 1/1

65 c

Vitale param.:

HF 147

ST-II -2,1

ST-aVF -2,6

ST-V4 0,0

VES 0

ST-III -3,1

ST-V1 -0,8

ST-V5 0,1

I.V ADENOSIN 6 mg

STindx 2,8

ST-aVR 0,6




ST-V2 -0,1

ST-V6 -0,1

ST-I 1,0


ST-aVL 2,0

ST-V3 -0,1



Case 66


Ubekræftet



25mm/s 10mm/mV 150Hz 9.0.10 Vogn-ID: 65535

EID: EDT: BESTILLING:

t = 1 time after debut




Udskrevet: 10, august, 2022 08:22:24 KIBI Report Generator V2.0.0 - Kibi Powered by Amps llc

Page 1 of 2

t = 1 døgn efter debut


Hvorfor ses ST-elevation ved
STEMI?

Case 67




Case 68

68 a




68 b




Case 69


VT 22-02-2018 19:12:04 Rytme : 194 bpm Varighed : 6,9 s QRS : 23




VT 22-02-2018 19:13:40 Rytme : 194 bpm Varighed : 1,3 s QRS : 5



VT 22-02-2018 19:11:46 Rytme : 200 bpm Varighed : 3,3 s QRS : 12




Ventrikulære triplets 22-02-2018 17:01:35 Varighed : 710 ms





Case 70




70 a



70 b



Case 71



05-02-2022 (lø) 07:57:17 Rapporteret anfald - fortsat

25 mm/s (Skala: 1:2)



05-02-2022 (lø) 07:57:32 Rapporteret anfald - fortsat

25 mm/s (Skala: 1:2)



05-02-2022 (lø) 07:57:48 Rapporteret anfald - fortsat



25 mm/s (Skala: 1:2)



Case 72

25 mm/s 10 mm/mV


HR 72 /min




Intervaller
RR 828 ms
P 120 ms
PR 132 ms
QRS 126 ms
QT 398 ms
QTc 442 ms


Aksen
P 64 °
QRS 28 °
T 106 °

Holter









25 mm/s

10,0 mm/mV

0.05-150 Hz

50 Hz

SBS

SSF

12-12-2012

01-10

Case 73

HR : 65 bpm
P : 120 ms
PR : 145 ms
QRS : 183 ms
QT/QTc : 442/460 ms
P/QRS/T : 44/41/178 deg.
RV5/SV1 : 1.441/1.606 mV


Diagnosis:
Sinus Rhythm
CLBBB(Complete Left Bundle Branch Block)(I,aVL,V1,V5,V6)

Report Confirmed By:



Case 74

02:02




GE CASE V6.5(0)

25mm/s 10mm/mV 0.01-100Hz 60Hz Spline

Unconfirmed

Primary physician

Page 1



GE CASE V6, 5 (0)
25mm/s 10mm/mV 0.01-100Hz 60Hz Spine

Unconfirmed

Primary
physician